

TIBC/% Saturation Iron Profile Analyzer

Order Name: **AN %SAT/TI**

Test Number: 4501525

Revision Date: 01/16/2024

TEST NAME	METHODOLOGY	LOINC CODE
Total Iron Binding Capacity		2500-7
Iron	Ferene	2498-4
% Saturation TIBC		2502-3
UIBC	Ferene	2501-5

SPECIMEN REQUIREMENTS				
Specimen	Specimen Volume (min)	Specimen Type	Specimen Container	Transport Environment
Preferred	1 mL (0.5 mL)	Serum	Clot Activator SST (Gold Top)	Refrigerated
Instructions	Specimen stability: Ambient 10 hours, Refrigerated 7 days, Frozen 1 year			

GENERAL INFORMATION	
Testing Schedule	Daily
Expected TAT	1-2 days
Notes	For ongoing iron determinations on a patient. The sample should be obtained at approximately the same time each day. The concentration may vary 30% throughout the day. If the Iron Level is less than 10 ug/dL then Estimation Iron with Total Iron Binding Capacity is ordered and resulted at no charge. Order Code: [0691091] .Iron with Total Iron Binding Capacity Result Code: [4501101] Estimated TIBC (LOINC 2500-7) Result Code: [4501151] Estimated % Saturation (LOINC 2502-3)
CPT Code(s)	83550, 83540
Lab Section	Chemistry